


**CHANNEL CALIBRATION PROCEDURE  
FOR THE NEES@UCSB WILDLIFE LIQUEFACTION  
ARRAY (WLA) AND 5210 SITE**

<b>CUSTOMER</b>	University Of California Santa Barbara Crustal Studies Department
<b>SYSTEM LOCATION</b>	WLA
<b>SYSTEM S/N</b>	207
<b>DATE OF TEST</b>	2/18/2012
<b>PERFORMED BY (Print)</b>	Robin Gee and Dan Radulescu
<b>SIGNATURE</b>	

# CHANNEL CALIBRATION PROCEDURE FOR THE WLA and 5210 SURFACE ARRAY

## 1.0 PURPOSE

The purpose of this procedure is the determination of the calibration factors for the entire system as described in the proposal No: DCR2006-001. The main components of the system will be checked for functionality and when needed a calibration factor will be determined. The sensors and the entire system shall be tested such that they respond within a specified range and accuracy to an input traceable to the National Bureau of Standards or an acceptable physical constant, (e.g., tilt testing of an accelerometer within the earth's gravitational field). This calibration will require temporary removal of the sensors from their normal location.

## 2.0 REFERENCES

- Kinematics Episensor (Model FBA ES-T) Guide
- WLA Instrumentation Guide ([nees.ucsb.edu/facilities/wla](http://nees.ucsb.edu/facilities/wla))

It is recommended that this calibration be performed every 12 months

## 3.0 SYSTEM PERFORMANCE NOTES

- Because this procedure is intended to be used by a qualified person, step-by-step instructions are not given
- Test sequence may be changed as needed for safety and /or efficiency.
- Items for which quantitative measurements cannot or need not to be made shall be reported in a qualitative mode (e.g. Yes/No).
- Any activities performed outside the normal scope of this procedure shall be documented.

- When a deficiency is observed, the technician may undertake additional testing and install factory authorized and/or factory calibrated replacement parts to restore the proper operation of the instrument.
- Calibration readings are equally valid using either the internal batteries (>11.5 VDC under load) or using an external power supply (between 12.0 and 13.0 VDC).

#### 4.0 TEST EQUIPMENT

Instrument Type	Manufacturer	Model	Range
Digital Voltmeter	FLUKE	189	2Vdc & 20Vdc
Bubble Level	PRO Products	PRO-inclinometer	0° ÷ 180°
Tilt Table	Radu Project	TT-1	±180°

Note: inclinometer accuracy to  $\pm 1/2^\circ$  (used to level tilt table)

#### 5.0 PRE-TEST CONDITIONS

- Notify the End user that the system will be taken out of normal operation conditions

(Initials) RB

- Check the overall system functionality and appearance. Document any observed anomaly. If a subassembly is not functional, document the findings, perform the repair first (if possible), and continue with the calibration

(Initials) RB

NOTES: System functional  
Sensor power supply functional  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

(Initials) RB

## 6.0 SYSTEM TEST

### 6.1 UNINTERRUPTIBLE POWER SUPPLY

- a) Check the battery charging Indicator.  
Mark FULL or indicate in % \_\_\_\_\_ (Initials) \_\_\_\_\_
- b) Disconnect the AC power cord and wait 10 minutes.  
The intermittent Battery operation sound should be present.  
The battery charging indicator shall stay on the same range.  
\_\_\_\_\_  
(Initials) \_\_\_\_\_
- c) Reconnect the AC \_\_\_\_\_ (Initials) \_\_\_\_\_
- d) Document when the battery has been installed (dd/mm/yy) \_\_\_\_\_

NOTES: Used 12v dc power supply

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

### 6.2 SENSOR CALIBRATION

#### WLA SURFACE ACCELEROMETERS

LOCATION CODE #	SENSOR MODEL	SENSOR S/N	CALIBRATION [V/g]	
			AS FOUND	AS LEFT
00	Triaxial Accelerometer X-axis	1457	9.96	9.96
00	Triaxial Accelerometer Y-axis	1457	9.95	9.95
00	Triaxial Accelerometer Z-axis	1457	9.88	9.88
10	Triaxial Accelerometer X-axis	1456	9.97	9.97
10	Triaxial Accelerometer Y-axis	1456	9.96	9.96
10	Triaxial Accelerometer Z-axis	1456	9.61	9.61
11	Triaxial Accelerometer X-axis	1458	9.97	9.97
11	Triaxial Accelerometer Y-axis	1458	9.96	9.96
11	Triaxial Accelerometer Z-axis	1458	9.84	9.84

NOTES: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**5210 SURFACE ACCELEROMETERS**

LOCATION CODE #	SENSOR MODEL	SENSOR S/N	CALIBRATION [V/g]	
			AS FOUND	AS LEFT
00	Triaxial Accelerometer X-axis	2075	9.98	9.98
00	Triaxial Accelerometer Y-axis	2075	9.96	9.96
00	Triaxial Accelerometer Z-axis	2075	9.98	9.98

NOTES: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**7.0 FINAL STEPS**

- Change the batteries from the UPS and SENSOR POWER SUPPLY if they are three (3) years old or more. If the batteries are not purchased, make a note and change them at the first maintenance visit.

(Initials) RB

- Return the system to functional state

(Initials) RB

- Attach Final record to this document

(Initials) RB

- List all test equipment

(Initials) RB

- Inform the end user that the system is functional

(Initials) RB

- Prepare the site (close the hat and the equipment)

(Initials) RB

NOTES: System functional  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**8.0 SUMMARY (Comments, Parts replaced, Deficiencies, etc.)**

System left functional  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**9.0 CERTIFICATION**

All items included in this procedure have been performed unless noted above and were found or have been adjusted to be within the range required by this procedure.

(yes/no) Yes  
\_\_\_\_\_  
Robin Gee  
(Signature)

Robin Gee  
(Print)

**10.0 ACTION REQUIRED (IF ANY)**

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

## **APPENDIX A**

Correspondence between the channel number and sensor type

Table A1: WLA SURFACE ARRAY INSTRUMENTS

<b>Location Code #</b>	<b>Sensor Type</b>	<b>Model / Manufacturer</b>
00	Triaxial Accelerometer X,Y,Z-axis	Kinematics FBA ES-T
10	Triaxial Accelerometer X,Y,Z-axis	Kinematics FBA ES-T
11	Triaxial Accelerometer X,Y,Z-axis	Kinematics FBA ES-T

Table A2: 5210 SURFACE ARRAY INSTRUMENTS

<b>Location Code #</b>	<b>Sensor Type</b>	<b>Model / Manufacturer</b>
00	Triaxial Accelerometer X,Y,Z-axis	Kinematics FBA ES-T

**NOTE:**

- Connected to the A/D Input through a Signal Conditioning board Model 163MK manufactured by CALEX



## **APPENDIX B**

### Recommended Calibration Methods

## ACCELEROMETERS

Each unit (uniaxial or triaxial that can be removed temporarily for calibration will be calibrated using a tilt table).

### For the horizontal axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER

WLA S00 SN 1457

The following readings shall be taken in the indicated order: X-AXIS

- |                            |                   |
|----------------------------|-------------------|
| 1. Horizontal              | <u>-0.242</u> [V] |
| 2. tilt to +30 degrees     | <u>-5.19</u> [V]  |
| 3. tilt to +90 degrees     | <u>-10.10</u> [V] |
| 4. tilt back to horizontal | <u>-0.239</u> [V] |
| 5. tilt to -30 degrees     | <u>4.75</u> [V]   |
| 6. tilt to -90 degrees     | <u>9.83</u> [V]   |
| 7. tilt back to horizontal | <u>-0.240</u> [V] |

### CALIBRATION FACTOR

$\{(|\text{Read3} - \text{Read1}|) + (|\text{Read6} - \text{Read4}|)\} / 2$

9.96 [V/g]

### For the vertical axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER

WLA S00 SN 1457

The following readings shall be taken in the indicated order: Z-AXIS

- |                           |                   |
|---------------------------|-------------------|
| 1. Horizontal             | <u>0.079</u> [V]  |
| 2. tilt to +60 degrees    | <u>-4.81</u> [V]  |
| 3. tilt to +90 degrees    | <u>-9.80</u> [V]  |
| 4. Continue to Horizontal | <u>-19.86</u> [V] |

### CALIBRATION FACTOR

$|\text{Read 3} - \text{Read 1}|$

9.88 [V/g]

# ACCELEROMETERS

Each unit (uniaxial or triaxial that can be removed temporarily for calibration will be calibrated using a tilt table).

## For the horizontal axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER

WLA 800 SN 1457

The following readings shall be taken in the indicated order:

Y-AXIS

- |                            |                   |
|----------------------------|-------------------|
| 1. Horizontal              | <u>-0.104</u> [V] |
| 2. tilt to +30 degrees     | <u>-4.89</u> [V]  |
| 3. tilt to +90 degrees     | <u>-10.07</u> [V] |
| 4. tilt back to horizontal | <u>0.102</u> [V]  |
| 5. tilt to -30 degrees     | <u>5.05</u> [V]   |
| 6. tilt to -90 degrees     | <u>9.82</u> [V]   |
| 7. tilt back to horizontal | <u>0.098</u> [V]  |

## CALIBRATION FACTOR

{(|Read3 – Read1|) + (|Read6 – Read4|)} / 2

9.95 [V/g]

## For the vertical axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER

The following readings shall be taken in the indicated order:

- |                           |           |
|---------------------------|-----------|
| 1. Horizontal             | _____ [V] |
| 2. tilt to +60 degrees    | _____ [V] |
| 3. tilt to +90 degrees    | _____ [V] |
| 4. Continue to Horizontal | _____ [V] |

## CALIBRATION FACTOR

|Read 3 – Read 1|

\_\_\_\_\_ [V/g]

## ACCELEROMETERS

Each unit (uniaxial or triaxial that can be removed temporarily for calibration will be calibrated using a tilt table).

### For the horizontal axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER

WLA 10 SN 1456

The following readings shall be taken in the indicated order:

X-AXIS

- |                            |                   |
|----------------------------|-------------------|
| 1. Horizontal              | <u>-0.139</u> [V] |
| 2. tilt to +30 degrees     | <u>-5.13</u> [V]  |
| 3. tilt to +90 degrees     | <u>-10.22</u> [V] |
| 4. tilt back to horizontal | <u>-0.138</u> [V] |
| 5. tilt to -30 degrees     | <u>4.84</u> [V]   |
| 6. tilt to -90 degrees     | <u>9.73</u> [V]   |
| 7. tilt back to horizontal | <u>-0.138</u> [V] |

### CALIBRATION FACTOR

$\{(|\text{Read3} - \text{Read1}|) + (|\text{Read6} - \text{Read4}|)\} / 2$

9.97 [V/g]

### For the vertical axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER

WLA 10 SN 1456

The following readings shall be taken in the indicated order:

Z-AXIS

- |                           |                   |
|---------------------------|-------------------|
| 1. Horizontal             | <u>-0.726</u> [V] |
| 2. tilt to +60 degrees    | <u>-5.69</u> [V]  |
| 3. tilt to +90 degrees    | <u>-10.37</u> [V] |
| 4. Continue to Horizontal | <u>-20.68</u> [V] |

### CALIBRATION FACTOR

$|\text{Read 3} - \text{Read 1}|$

9.61 [V/g]

## v ACCELEROMETERS

Each unit (uniaxial or triaxial that can be removed temporarily for calibration will be calibrated using a tilt table).

### For the horizontal axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER

WLA 10 SN 1456

The following readings shall be taken in the indicated order:

Y-AXIS

- |                            |             |
|----------------------------|-------------|
| 1. Horizontal              | - 0.009 [V] |
| 2. tilt to +30 degrees     | - 4.99 [V]  |
| 3. tilt to +90 degrees     | - 10.02 [V] |
| 4. tilt back to horizontal | - 0.012 [V] |
| 5. tilt to -30 degrees     | 4.96 [V]    |
| 6. tilt to -90 degrees     | 0.90 [V]    |
| 7. tilt back to horizontal | 0.013 [V]   |

### CALIBRATION FACTOR

$\{(|\text{Read3} - \text{Read1}|) + (|\text{Read6} - \text{Read4}|)\} / 2$

9.96 [V/g]

### For the vertical axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER

The following readings shall be taken in the indicated order:

- |                           |           |
|---------------------------|-----------|
| 1. Horizontal             | _____ [V] |
| 2. tilt to +60 degrees    | _____ [V] |
| 3. tilt to +90 degrees    | _____ [V] |
| 4. Continue to Horizontal | _____ [V] |

### CALIBRATION FACTOR

$|\text{Read 3} - \text{Read 1}|$

\_\_\_\_\_ [V/g]

## ACCELEROMETERS

Each unit (uniaxial or triaxial that can be removed temporarily for calibration will be calibrated using a tilt table).

### For the horizontal axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER

WLA 11 SN 1458

The following readings shall be taken in the indicated order:

X-AXIS

1. Horizontal	<u>-0.305</u> [V]
2. tilt to +30 degrees	<u>-5.27</u> [V]
3. tilt to +90 degrees	<u>-10.19</u> [V]
4. tilt back to horizontal	<u>-0.326</u> [V]
5. tilt to -30 degrees	<u>4.67</u> [V]
6. tilt to -90 degrees	<u>9.73</u> [V]
7. tilt back to horizontal	<u>0.327</u> [V]
<b>CALIBRATION FACTOR</b> {( Read3 – Read1 ) + ( Read6 – Read4 )} / 2	<u>9.97</u> [V/g]

### For the vertical axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER

WLA 11 SN 1458

The following readings shall be taken in the indicated order:

Z-AXIS

1. Horizontal	<u>-0.281</u> [V]
2. tilt to +60 degrees	<u>-5.36</u> [V]
3. tilt to +90 degrees	<u>-10.12</u> [V]
4. Continue to Horizontal	<u>-20.24</u> [V]
<b>CALIBRATION FACTOR</b>  Read 3 – Read 1	<u>9.84</u> [V/g]

## ACCELEROMETERS

Each unit (uniaxial or triaxial that can be removed temporarily for calibration will be calibrated using a tilt table).

### For the horizontal axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER

WLA 11 SN 1458

The following readings shall be taken in the indicated order:

Y-AXIS

- |                            |                   |
|----------------------------|-------------------|
| 1. Horizontal              | <u>-0.291</u> [V] |
| 2. tilt to +30 degrees     | <u>-5.28</u> [V]  |
| 3. tilt to +90 degrees     | <u>-10.29</u> [V] |
| 4. tilt back to horizontal | <u>-0.324</u> [V] |
| 5. tilt to -30 degrees     | <u>4.63</u> [V]   |
| 6. tilt to -90 degrees     | <u>9.59</u> [V]   |
| 7. tilt back to horizontal | <u>-0.329</u> [V] |

### CALIBRATION FACTOR

$\{(|\text{Read3} - \text{Read1}|) + (|\text{Read6} - \text{Read4}|)\} / 2$

9.96 [V/g]

### For the vertical axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER \_\_\_\_\_

The following readings shall be taken in the indicated order:

- |                           |           |
|---------------------------|-----------|
| 1. Horizontal             | _____ [V] |
| 2. tilt to +60 degrees    | _____ [V] |
| 3. tilt to +90 degrees    | _____ [V] |
| 4. Continue to Horizontal | _____ [V] |

### CALIBRATION FACTOR

$|\text{Read 3} - \text{Read 1}|$

\_\_\_\_\_ [V/g]

## ACCELEROMETERS

Each unit (uniaxial or triaxial that can be removed temporarily for calibration will be calibrated using a tilt table).

### For the horizontal axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER

5210 S00 SN 2075

The following readings shall be taken in the indicated order:

X - AXIS

- |                            |                   |
|----------------------------|-------------------|
| 1. Horizontal              | <u>-0.128</u> [V] |
| 2. tilt to +30 degrees     | <u>-5.21</u> [V]  |
| 3. tilt to +90 degrees     | <u>-10.25</u> [V] |
| 4. tilt back to horizontal | <u>-0.129</u> [V] |
| 5. tilt to -30 degrees     | <u>4.84</u> [V]   |
| 6. tilt to -90 degrees     | <u>9.71</u> [V]   |
| 7. tilt back to horizontal | <u>0.137</u> [V]  |

### CALIBRATION FACTOR

$\{(|\text{Read3} - \text{Read1}|) + (|\text{Read6} - \text{Read4}|)\} / 2$

9.98 [V/g]

### For the vertical axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER

5210 S00 SN 2075

The following readings shall be taken in the indicated order:

Z - AXIS

- |                           |                   |
|---------------------------|-------------------|
| 1. Horizontal             | <u>-0.035</u> [V] |
| 2. tilt to +60 degrees    | <u>-4.96</u> [V]  |
| 3. tilt to +90 degrees    | <u>-10.01</u> [V] |
| 4. Continue to Horizontal | <u>-20.02</u> [V] |

### CALIBRATION FACTOR

$|\text{Read 3} - \text{Read 1}|$

9.98 [V/g]



# ACCELEROMETERS

Each unit (uniaxial or triaxial that can be removed temporarily for calibration will be calibrated using a tilt table).

## For the horizontal axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER

5210 SØØ SN 2075

The following readings shall be taken in the indicated order:

Y-AXIS

- |                            |                   |
|----------------------------|-------------------|
| 1. Horizontal              | <u>-0.306</u> [V] |
| 2. tilt to +30 degrees     | <u>-5.29</u> [V]  |
| 3. tilt to +90 degrees     | <u>-10.33</u> [V] |
| 4. tilt back to horizontal | <u>-0.303</u> [V] |
| 5. tilt to -30 degrees     | <u>4.65</u> [V]   |
| 6. tilt to -90 degrees     | <u>9.59</u> [V]   |
| 7. tilt back to horizontal | <u>-0.311</u> [V] |

## CALIBRATION FACTOR

$\{(|\text{Read3} - \text{Read1}|) + (|\text{Read6} - \text{Read4}|)\} / 2$

9.96 [V/g]

## For the vertical axis:

The unit shall be placed on a tilt table which has been leveled and checked with a mechanical angular device.

MODEL / SERIAL NUMBER

\_\_\_\_\_

The following readings shall be taken in the indicated order:

- |                           |           |
|---------------------------|-----------|
| 1. Horizontal             | _____ [V] |
| 2. tilt to +60 degrees    | _____ [V] |
| 3. tilt to +90 degrees    | _____ [V] |
| 4. Continue to Horizontal | _____ [V] |

## CALIBRATION FACTOR

$|\text{Read 3} - \text{Read 1}|$

\_\_\_\_\_ [V/g]



MICRO PRECISION CALIBRATION INC.  
 9080 ACTIVITY RD, STE C  
 SAN DIEGO  
 (858) 547-0217

## Certificate of Calibration

Date: 6/17/2011

Certificate #: 1404392

**Customer:**

ENGLEKIRK STRUCTURAL ENGINEERING CENTER  
 10201 POMERADO RD  
 SAN DIEGO, CA, 92131

Work Order: 120733

MPC Control #: BL8539  
 Asset ID: N/A  
 Gage Type: TILT TABLE  
 Manufacturer: KINEMATRICS  
 Model Number: F100200-XX-PL  
 Size: N/A  
 Temp./RH: 21 °C / 55 %

Serial Number: 172  
 Department: N/A  
 Performed By: PETER SINKS  
 Received Condition: IN TOLERANCE  
 Returned Condition: IN TOLERANCE  
 Cal Date: June 17, 2011  
 Cal. Interval: 12 MONTHS  
 Cal. Due Date: June 17, 2012

**Found conditions meet or exceed manufacturer specifications.**

\*Calibration Notes:

**Test Points**

Description	Standard	Tolerance -	Tolerance +	As Found	As Left	UOM	Result
TILT DEVIATION	-30.00	-31.00	-29.00	-29.95	-29.95	Deg	Pass
	-60.00	-61.00	-59.00	-60.00	-60.00	Deg	Pass
	-90.00	-91.00	-39.00	-89.95	-89.95	Deg	Pass
	30.00	29.00	31.00	30.05	30.05	Deg	Pass
	60.00	59.00	61.00	60.05	60.05	Deg	Pass
	90.00	89.00	91.00	90.00	90.00	Deg	Pass

**Standards Used To Calibrate Equipment**

I.D.	Description	Model	Serial	Manufacturer	Cal. Due Date	Traceability #
B3864	PROTRACTOR	N/A	N/A	HELIOS	4/30/2012	1327219
T9186	HEIGHT GAGE	0 - 12 IN	1886	VERDICT	6/30/2011	1005841

Calibrating Technician:

*Peter Sinks*

PETER SINKS

QC Approval:

*Maurice Heath*

MAURICE HEATH

Unless Otherwise Noted, Uncertainty Estimated at >= 4 to 1. Uncertainties have been estimated at a 95 percent confidence level (k=2) Services rendered comply with ISO 17025:2005, ISO 9001:2008, ANSINC/SLS Z540-3, MPC Quality Manual, MPC CSD, and with customer purchase order instructions.

Calibration cycles and resulting due dates were submitted/approved by the customer. Any number of factors may cause an instrument to drift out of tolerance before the next scheduled calibration. Recalibration cycles should be based on frequency of use, environmental conditions and customer's established systematic accuracy. The information on this report, pertains only to the instrument identified.

All standards are traceable to the National Institute of Standards and Technology (NIST). Services rendered include proper manufacturer's service instructions and are warranted for no less than thirty (30) days. This report may not be reproduced in part or in whole without the prior written approval of the issuing MPC Lab.



MICRO PRECISION CALIBRATION INC.  
 9080 ACTIVITY RD, STE C  
 SAN DIEGO  
 (858) 547-0217

## Certificate of Calibration

Date: 6/17/2011

Certificate #: 1404392

### Procedures Used In This Event:

Procedure Name

Description

ANGLE PLATE

ANGLE PLATES - PROC# CP00206

Calibrating Technician:

*Peter Sinks*

PETER SINKS

QC Approval:

*Maurice Heath*

MAURICE HEATH

Unless Otherwise Noted, Uncertainty Estimated at  $\geq 4$  to 1. Uncertainties have been estimated at a 95 percent confidence level (k=2). Services rendered comply with ISO 17025:2005, ISO 9001:2008, ANSI/NCCL Z540-3, MPC Quality Manual, MPC CSD and with customer purchase order instructions.

Calibration cycles and resulting due dates were submitted/approved by the customer. Any number of factors may cause an instrument to drift out of tolerance before the next scheduled calibration. Recalibration cycles should be based on frequency of use, environmental conditions and customer's established systematic accuracy. The information on this report, pertains only to the instrument identified.

All standards are traceable to the National Institute of Standards and Technology (NIST). Services rendered include proper manufacturer's service instructions and are warranted for no less than thirty (30) days. This report may not be reproduced in part or in whole without the prior written approval of the issuing MPC lab.



MICRO PRECISION CALIBRATION INC.  
 9080 ACTIVITY RD, STE C  
 SAN DIEGO  
 (858) 547-0217



# Certificate of Calibration

Date: 12/19/2011

Lab # 935.10

Certificate #: 1620009

**Customer:**

ENGLEKIRK STRUCTURAL ENGINEERING CENTER  
 10201 POMERADO RD  
 SAN DIEGO, CA, 92131

Work Order: 134240

MPC Control #: BL3668  
 Asset ID: N/A  
 Gage Type: DIGITAL MULTIMETER  
 Manufacturer: FLUKE  
 Model Number: 85 III  
 Size: N/A  
 Temp./RH: 19.4 °C / 47 %

Serial Number: 46450375  
 Department: N/A  
 Performed By: FRANCISCO OLIVEROS  
 Received Condition: IN TOLERANCE  
 Returned Condition: IN TOLERANCE  
 Cal Date: December 12, 2011  
 Cal. Interval: 12 MONTHS  
 Cal. Due Date: December 12, 2012

**Found conditions meet or exceed manufacturer specifications.**

**\*Calibration Notes:**

Data and uncert. attached.

**Standards Used To Calibrate Equipment**

I.D.	Description	Model	Serial	Manufacturer	Cal. Due Date	Traceability #
BC3293	METER CALIBRATOR	5700A	5990310	FLUKE	8/31/2012	1254320
T9825	UNIVERSAL CALIBRATOR	9100	40144	WAVETEK	4/30/2012	1342582

**Procedures Used In This Event:**

Procedure Name	Description
MULTIMETER (FLUKE) GENE	MULTIMETER, DIG 33K8-4-14-1 (FLUKE)

Calibrating Technician:

FRANCISCO OLIVEROS


QC Approval:

Ricardo Morris

Unless Otherwise Noted, Uncertainty Estimated at  $\geq 4$  to 1. Uncertainties have been estimated at a 95 percent confidence level (k=2). Services rendered comply with ISO 17025:2005, ISO 9001:2008, ANSINCSSL Z540-3, MPC Quality Manual, MPC CSD and with customer purchase order instructions.

Calibration cycles and resulting due dates were submitted/approved by the customer. Any number of factors may cause an instrument to drift out of tolerance before the next scheduled calibration. Recalibration cycles should be based on frequency of use, environmental conditions and customer's established systematic accuracy. The information on this report, pertains only to the instrument identified.


All standards are traceable to the National Institute of Standards and Technology (NIST). Services rendered include proper manufacturer's service instructions and are warranted for no less than thirty (30) days. This report may not be reproduced in part or in whole without the prior written approval of the issuing MPC lab.



CUSTOMER	ENGLEKIRK	MFR	FLUKE
TECH #	FRANCISCO O.	MODEL #	85III
DATE	12/12/2011	MCP #	BL3668
ASSET #	N/A	WORK #	134240
SERIAL #	46450375		

STEP	FUNCTION TESTED	NOMINAL VALUE	VALUE AS RECEIVED	RETURNED VALUE	OUT-OF-TOL	TOLERANCE BAND
4.1	DC Volts					
	400 mV Range	350.0 mV	349.1	349.1		349.5 to 350.5 mV
	4.0 V Range	3.500 V	3.498	3.498		3.495 to 3.505 V
	40.0 V Range	35.00 V	34.98	34.98		34.95 to 35.05 V
	40.0 V Range	-35.00 V	-34.98	-34.98		-35.05 to -34.95 V
	400.0 V Range	350.0 V	349.8	349.8		349.5 to 350.5 V
	1000 V Range	1000 V	1000	1000		998 to 1002 V
4.2	AC Volts					
	400 mV Range @ 60 Hz	350.0 mV	349.1	349.1		347.8 to 352.2 mV
	400 mV Range @ 5 kHz	350.0 mV	350.0	350.0		346.1 to 353.9 mV
	400 mV Range @ 20 kHz	350.0 mV	350.1	350.1		342.6 to 357.4 mV
	4.0 V Range @ 60 Hz	3.500 V	3.497	3.497		3.480 to 3.520 V
	4.0 V Range @ 5 kHz	3.500 V	3.485	3.485		3.463 to 3.537 V
	4.0 V Range @ 20 kHz	3.500 V	3.497	3.497		3.356 to 3.644 V
	40 V Range @ 60 Hz	35.00 V	34.99	34.99		34.80 to 35.20 V
	40 V Range @ 5 kHz	35.00 V	34.98	34.98		34.63 to 35.37 V
	40 V Range @ 20 kHz	35.00 V	34.97	34.97		33.56 to 36.44 V
	400 V Range @ 60 Hz	350.0 V	349.7	349.7		348.0 to 352.0 V
	400 V Range @ 5 kHz	350.0 V	350.1	350.1		346.3 to 353.7 V
	400 V Range @ 20 kHz	100.0 V	98.9	98.9		95.6 to 104.4 V
	400 V Range @ 20 kHz	200.0 V	199.8	199.8		191.6 to 208.4 V
	400 V Range @ 20 kHz	300.0 V	300.2	300.2		287.6 to 312.4 V
	400 V Range @ 20 kHz	350.0 V	349.7	349.7		335.6 to 364.4 V
	1000 V Range @ 60 Hz	1000 V	999	999		993 to 1007 V
	1000 V Range @ 5 kHz	1000 V	998	998		978 to 1022 V
4.3	Ohms					
	400 Ω Range	190.0 Ω	190.3	190.3		189.5 to 190.5 Ω
	4.0 kΩ Range	1.900 kΩ	1.900	1.900		1.895 to 1.905 kΩ
	40 kΩ Range	19.00 kΩ	19.99	19.99		18.95 to 19.05 kΩ






CUSTOMER ENGLEKIRK  
 TECH # FRANCISCO O.  
 DATE 12/12/2011  
 ASSET # N/A  
 SERIAL # 46450375

MFR FLUKE  
 MODEL # 85III  
 MCP # BL3668  
 WORK # 134240

STEP	FUNCTION TESTED	NOMINAL VALUE	VALUE AS RECEIVED	RETURNED VALUE	OUT-OF-TOL	TOLERANCE BAND
4.3	Ohms	(Cont'd)				
	4 MΩ Range	1.900 MΩ	1.900	1.900		1.895 to 1.905 MΩ
	40 MΩ Range	10.00 MΩ	9.96	9.96		9.87 to 10.13 MΩ
	40 MΩ Range	19.00 MΩ	18.90	18.90		18.78 to 19.22 MΩ
	40 nS Range	10 nS (100.00 MΩ)	10.01	10.01		9.80 to 10.20 nS
4.4	Capacitance					
	5.0 μF Range	1.00 μF	.99	.99		0.97 to 1.03 μF
	0.5 μF Range	0.470 μF	0.470	0.470		0.463 to 0.477 μF
	0.05 μF Range	0.0470 μF	.0468	.0468		0.0463 to 0.0477 μF
	5 nF Range (Relative Mode)	4.70 nF	4.69	4.69		4.63 to 4.77 nF
4.5	mA ( DC )					
	400 μA Range	350.0 μA	350.0	350.0		349.1 to 350.9 μA
	4000 μA Range	3500 μA	3498	3498		3491 to 3509 μA
	40 mA Range	35.00 mA	34.99	34.99		34.91 to 35.09 mA
	400 mA Range	350.0 mA	349.7	349.7		349.1 to 350.9 mA
	4000 mA Range	3500 mA	3501	3501		3491 to 3509 mA
	10 A Range	10.00 A	10.00	10.00		9.96 to 10.04 A
4.6	mA ( AC )					
	400 μA Range @ 60 Hz	350.0 μA	349.6	349.6		347.7 to 352.3 μA
	400 μA Range @ 1kHz	350.0 μA	349.8	349.8		347.7 to 352.3 μA
	3500 μA Range @ 60 Hz	3500 μA	3506	3506		3477 to 3523 μA
	3500 μA Range @ 1kHz	3500 μA	3505	3505		3477 to 3523 μA
	40 mA Range @ 60 Hz	35.0 mA	34.97	34.97		34.77 to 35.23 mA
	40 mA Range @ 1kHz	35.0 mA	34.98	34.98		34.77 to 35.23 mA
	400 mA Range @ 60 Hz	350.0 mA	350.6	350.6		347.7 to 352.3 mA
	400 mA Range @ 1kHz	350.0 mA	350.4	350.4		347.7 to 352.3 mA
	4000 mA Range @ 60 Hz	3500 mA	3500	3500		3477 to 3523 mA



CUSTOMER ENGLEKIRK  
 TECH # FRANCISCO O.  
 DATE 12/12/2011  
 ASSET # N/A  
 SERIAL # 46450375

MFR FLUKE  
 MODEL # 85III  
 MCP # BL3668  
 WORK # 134240

STEP	FUNCTION TESTED	NOMINAL VALUE	VALUE AS RECEIVED	RETURNED VALUE	OUT-OF-TOL	TOLERANCE BAND
	4000 mA Range @ 1kHz	3500 mA	3498	3498		3477 to 3523 mA
4.6	mA ( AC )	(Cont'd)				
	10 A Range @ 60 Hz	10.00 A	10.0	10.0		9.92 to 10.08 A
	10 A Range @ 1kHz	10.00 A	9.99	9.99		9.92 to 10.08 A
4.7	Diode Test					
		3.000 V	2.997	2.997		2.939 to 3.061 V
4.8	Frequency Test					
	@ 150 mV	19.000 kHz	19.000	19.000		18.998 to 19.002 kHz
	@ 150 mV	190.00 kHz	190.00	190.00		189.98 to 190.02 kHz

COMMENTS:


# Uncertainty Analysis Report

12-Dec-11

## Englekirk

Submitted by: Francisco Oliveros

Approved By: Ricardo Morris

Asset: BL3668

Subject Unit: Multimeter

Manufacturer: Fluke Corp.

Model Number: 85 III

## Uncertainty Results:

Parameter Name	Uncertainty	Units	Confidence Level
Resistance-Range			
<190K	+/- 6	mΩ	95.00
>190K	+/- .6	mΩ	95.00
DC Voltage-Ranges			
200 mV	+/- 6	μV	95.00
2V	+/- 600	μV	95.00
20 V	+/- 6	mV	95.00
200 V	+/- 40	mV	95.00
AC Voltage			
200 mV	+/- 6	μV	95.00
2V	+/- 60	μV	95.00
20 V	+/- 600	μV	95.00
200 V	+/- 6	mV	95.00
DC Current			
200 mA	+/- 6	μA	95.00
2 A	+/- 600	μA	95.00
AC Current			
200 mA	+/- 6	μA	95.00
2 A	+/- 6	mA	95.00

The expanded uncertainty presented in this report is consistent with the 1993 ISO Guide to the Expression of Uncertainty in Measurement.

The expanded Uncertainty is not to be confused with a tolerance limit for the user during application.